

AV. AGUILA AZTECA 20051-2E
TIJUANA, BC, MX

**ON SITE
CERTIFICATE OF CALIBRATION
FOR
ITT MEXICO**

Description: **ASHCROFT, 0 to 200 psi, Pressure Gage**

Serial No: **NONE**

Asset No: **731**

Simco ID: **41829-115**

Dept: **MEXICO**

PO No: **795729**

Calibration Date: 07/12/07	Calibration Interval: 6 Months	Recall Date: 01/12/08
Arrival Condition: MEETS MANUFACTURER'S SPEC'S.		Service: CALIBRATED TO MFR SPEC,& CLEAN

Procedure: **635-0506 REV 2**

Temperature: **70°F**

Relative Humidity: **42%**

Standards Used:

Type	Simco ID	Due Date	Intvl Mos	Acc/Unc	Trace No.
Documenting Process Calibrator	43023*36	05/29/08	12	PRESSURE	1180446298
Pressure Module	43023*95	12/01/07	12	PRESSURE	1166617526

Calibration Data:

Parameter	Nominal	Measured Before	Measured After	Tolerance
PRESSURE	20.0 PSI	19.86	19.86	+/-0.5 PSI
PRESSURE	60.0 PSI	59.68	59.68	+/-0.5 PSI
PRESSURE	80.0 PSI	79.82	79.82	+/-0.5 PSI
PRESSURE	100.0 PSI	99.93	99.93	+/-0.5 PSI
PRESSURE	140.0 PSI	140.24	140.24	+/-0.5 PSI
PRESSURE	200.0 PSI	200.26	200.26	+/-0.5 PSI

Work performed by:

Eduardo Miguel
Elect Tech B (X6026)

Reviewed by:

Aileen J Respicio
Customer Service Rep B

SIMCO Electronics' quality management system conforms to ISO 9001:2000, ISO/IEC 17025:1999, and ANSI/NCSL Z540-1-1994. All calibrations are performed using internationally recognized standards traceable to the International System of Units (SI Units). Traceability is achieved through calibrations by the National Institute of Standards and Technology (NIST), other National Measurement Institutes (NMIs'), or by using natural physical constants, intrinsic standards or ratio calibration techniques. Instruments are calibrated with a test accuracy ratio of 4:1 or greater, otherwise measurement uncertainty analysis and/or guard bands are applied during the measurement process. The information shown on this certificate applies only to the instrument identified above and may not be reproduced, except in full, without prior written consent from SIMCO Electronics. There is no implied warranty that the instrument will maintain its specified tolerances during the calibration interval due to possible drift, environment, or other factors beyond our control.

Dated: **07/12/07**

